

**510(k) Summary**  
**Prepared February 3, 2009**

**Sponsor:** Siemens Medical Solutions, Inc.,  
Ultrasound Division  
1230 Shorebird Way  
Mountain View, California 94043

**FEB 20 2009**

**Contact Person:** Shelly Pearce  
Telephone: (650) 694-5988  
Fax: (650) 694-5580

**Submission Date:** February 3, 2009

**Device Name:** Acuson S2000™ Diagnostic Ultrasound System

**Common Name:** Diagnostic Ultrasound System

**Classification:**

Regulatory Class: II  
Review Category: Tier II  
Classification Panel: Radiology

Ultrasonic Pulsed Doppler Imaging System	FR # 892.1550	Product Code 90-IYN
Ultrasonic Pulsed Echo Imaging System	FR # 892.1560	Product Code 90-IYO
Diagnostic Ultrasound Transducer	FR # 892.1570	Product Code 90-ITX

**A. Legally Marketed Predicate Devices**

The Acuson S2000™ Ultrasound System is substantially equivalent to the Acuson Antares Ultrasound System.

**B. Device Description:**

The Acuson S2000™ has been designed to meet the following product safety standards:

- UL 60601-1, Safety Requirements for Medical Equipment
- IEC 60601-2-37 Diagnostic Ultrasound Safety Standards
- CSA C22.2 No. 601-1, Safety Requirements for Medical Equipment
- AIUM/NEMA UD-3, Standard for Real Time Display of Thermal and Mechanical Acoustic Output Indices on Diagnostic Ultrasound Equipment
- AIUM/NEMA UD-2, Acoustic Output Measurement Standard for Diagnostic Ultrasound
- 93/42/EEC Medical Devices Directive
- Safety and EMC Requirements for Medical Equipment
  - EN/IEC 60601-1
  - EN/IEC 60601-1-1
  - EN/IEC 60601-1-2
- IEC 1157 Declaration of Acoustic Power
- ISO 10993-1 Biocompatibility

### C. Intended Use

The S2000™ ultrasound imaging systems are intended for the following applications: Fetal, Abdominal, Intraoperative, Pediatric, Small Parts, Transcranial, OB/GYN, Cardiac, Pelvic, Neonatal/Adult Cephalic, Vascular, Musculoskeletal, Superficial Musculoskeletal, and Peripheral Vascular applications.

The system also provides the ability to measure anatomical structures {fetal, abdominal, intraoperative, intraoperative neurological, pediatric, small organ, neonatal cephalic, adult cephalic, cardiac, trans-esophageal, transrectal, transvaginal, peripheral vessel, musculo-skeletal (conventional), musculo-skeletal (superficial) and neonatal cardiac} and calculation packages that provide information that provide information to the clinician that may be used adjunctively with other medical data obtained by a physician for clinical diagnosis purposes.

The Arterial Health Package (AHP) software provides the physician with the capability to measure Intima Media Thickness and the option to reference normative tables that have been validated and published in peer-reviewed studies. The information is intended to provide the physician with an easily understood tool for communicating with patients regarding state of their cardiovascular system. This feature should be utilized according to the "ASE Consensus Statement; Use of Carotid Ultrasound to Identify Subclinical Vascular Disease and Evaluate Cardiovascular Disease Risk: A Consensus Statement from the American Association of Echocardiography; Carotid Intima-Media Thickness Task Force, Endorsed by the Society for Vascular Imaging".



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration  
9200 Corporate Boulevard  
Rockville MD 20850

Siemens Medical Solution USA, Inc., Ultrasound Group  
% Mr. Mark Job  
Responsible Third Party Official  
Regulatory Technology Services LLC  
1394 25<sup>th</sup> Street NW  
BUFFALO MN 55313

FEB 20 2009

Re: K090334

Trade/Device Name: ACUSON S2000™ Diagnostic Ultrasound System

Regulation Number: 21 CFR 892.1550

Regulation Name: Ultrasonic pulsed doppler imaging system

Regulatory Class: II

Product Code: IYN, IYO, and ITX

Dated: February 9, 2009

Received: February 10, 2009

Dear Mr. Job:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

This determination of substantial equivalence applies to the following transducers intended for use with the ACUSON S2000™ Diagnostic Ultrasound System, as described in your premarket notification:

Transducer Model Number

CW2 Probe  
CW5 Probe  
EC9-4 Curved Array  
9L4 Linear Array  
14L5 Multi-D Array

4P1 Phased Array  
6C2 Curved Array  
4C1 Curved Array  
4V1 Phased Array  
10V4 Phased Array

14L5 SP Linear Array  
7CF2 Curved Array  
9EVF4 Curved Array  
V5Ms Multiplane TEE  
17L5HDS Linear Array

18L6 HD Linear Array  
8V3 Phased Array  
4V1c Phased Array  
6L3  
EV8C4

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus permits your device to proceed to market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (240) 276-0120. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>

If you have any questions regarding the content of this letter, please contact Lauren Hefner at (240) 276-3666.

Sincerely yours,



Janine M. Morris  
Acting Director, Division of Reproductive,  
Abdominal, and Radiological Devices  
Office of Device Evaluation  
Center for Devices and Radiological Health

Enclosure(s)

### 1.3 Indications for Use

510(k) Number (if known): K090334

Device Name: S2000™ Diagnostic Ultrasound System

Indications for Use: K090334

The S2000™ ultrasound imaging systems are intended for the following applications: Fetal, Abdominal, Intraoperative, Pediatric, Small Parts, Transcranial, OB/GYN, Cardiac, Pelvic, Neonatal/Adult Cephalic, Vascular, Musculoskeletal, Superficial Musculoskeletal, and Peripheral Vascular applications.

The system also provides the ability to measure anatomical structures (fetal, abdominal, intraoperative, intraoperative neurological, pediatric, small organ, neonatal cephalic, adult cephalic, cardiac, trans-esophageal, transrectal, transvaginal, peripheral vessel, musculo-skeletal (conventional), musculo-skeletal (superficial) and neonatal cardiac) and calculation packages that provide information that provide information to the clinician that may be used adjunctively with other medical data obtained by a physician for clinical diagnosis purposes.

The Arterial Health Package (AHP) software provides the physician with the capability to measure Intima Media Thickness and the option to reference normative tables that have been validated and published in peer-reviewed studies. The information is intended to provide the physician with an easily understood tool for communicating with patients regarding state of their cardiovascular system. This feature should be utilized according to the "ASE Consensus Statement; Use of Carotid Ultrasound to Identify Subclinical Vascular Disease and Evaluate Cardiovascular Disease Risk: A Consensus Statement from the American Association of Echocardiography; Carotid Intima-Media Thickness Task Force, Endorsed by the Society for Vascular Imaging".

Prescription Use X  
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use \_\_\_\_\_  
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

  
(Division Sign-Off)

Division of Reproductive, Abdominal and  
Radiological Devices

510(k) Number

K090334

### 1.3 Indications for Use Forms

#### Diagnostic Ultrasound Indications for Use Form

510 (k) Number (if known):

Device Name:

ACUSON S2000 Ultrasound System


Intended Use:

Ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation									
	A	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Color Velocity Imaging	Combined (Specify)	Other (Specify)
Ophthalmic										
Fetal		P	P	P	P	P	P		BMDC	Note 2,3,4,5,7,8,10, 11, 13
Abdominal		P	P	P	P	P	P		BMDC	Note 2,3,4,5,7,8,10, 11, 13
Intraoperative (Note 9)		P	P	P	P	P	P		BMDC	Note 2,3,4,5,7,8,10, 11, 14
Intraoperative Neurological		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11, 14
Pediatric		P	P	P	P	P	P		BMDC	Note 2,3,4,5,7,8,10, 11
Small Organ (Note 1)		P	P	P	P	P	P		BMDC	Note 2,3,4,5,7,8,10, 11, 14
Neonatal Cephalic		P	P	P	P	P	P		BMDC	Note 2,3,4,5,7,8,10
Adult Cephalic		P	P	P	P	P	P		BMDC	Note 2,3,4,5,7,8,10
Cardiac		P	P	P	P	P	P		BMDC	Note 2,3,4,5,6,7,8,10,15
Trans-esophageal		P	P	P	P	P	P		BMDC	
Transrectal		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11, 14
Transvaginal		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11
Transurethral										
Intravascular										
Peripheral vessel		P	P	P	P	P	P		BMDC	Note 2,3,4,5,6,7,8,10, 11, 14, 15
Laparoscopic										
Musculo-skeletal Conventional		P	P	P	P	P	P		BMDC	Note 2,3,4,5,7,8,10, 11, 14
Musculo-skeletal Superficial		P	P	P	P	P	P		BMDC	Note 2,3,4,5,7,8,10, 11, 14
Other (specify) Neonatal Cardiac		P	P	P	P	P	P		BMDC	Note 3,4,6

N = new indication; P = previously cleared by FDA, K072786

- Note 1 For example: breast, testes, thyroid, penis, prostate, etc.  
 Note 2 Ensemble tissue harmonic imaging  
 Note 3 SieClear multi-view spatial compounding  
 Note 4 Tissue Equalization Technology  
 Note 5 3-Scape real-time 3D imaging  
 Note 6 Cadence contrast agent imaging  
 Note 7 B&W SieScape panoramic imaging  
 Note 8 Power SieScape panoramic imaging  
 Note 9 For example: vascular, abdominal  
 Note 10 Clarify VE vascular enhancement technology  
 Note 11 Advanced Sieclear spatial compounding  
 Note 13 STIC  
 Note 14 eSie™ Touch elasticity imaging

Note 15   
 (Division Sign-Off)  
 Division of Reproductive, Abdominal and  
 Radiological Devices  
 510(k) Number K090334

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use (Per 21 CFR 801.10)

### Diagnostic Ultrasound Indications for Use Form

510 (k) Number (if known):

Device Name:

CW2 Probe for use with ACUSON S2000

Intended Use:

Ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation									
	A	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Color Velocity Imaging	Combined (Specify)	Other (Specify)
Ophthalmic										
Fetal					P					
Abdominal					F					
Intraoperative (Note 9)					P					
Intraoperative Neurological										
Pediatric					P					
Small Organ (Note 1)					P					
Neonatal Cephalic					P					
Adult Cephalic					P					
Cardiac					P					
Trans-esophageal										
Transrectal										
Transvaginal										
Transurethral										
Intravascular										
Peripheral vessel					P					
Laparoscopic										
Musculo-skeletal Conventional					P					
Musculo-skeletal Superficial					P					
Other (specify)										

Note: new indication, P = previously cleared by FDA # DE 803, K072705

Additional Comments:

Note 1 For example: breast, testes, thyroid, penis, prostate, etc.

Note 9 For example: vascular, abdominal

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use (Per 21 CFR 801.109)

(Division Sign-Off)

Division of Reproductive, Abdominal and  
Radiological Devices

510(k) Number

K090334

Diagnostic Ultrasound Indications for Use Form

510 (k) Number (if known):

Device Name:

CW5 Probe for use with ACUSON S2000

Intended Use:

Ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation									
	A	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Color Velocity Imaging	Combined (Specify)	Other (Specify)
Ophthalmic										
Fetal					P					
Abdominal					P					
Intraoperative (Note 9)					P					
Intraoperative Neurological										
Pediatric					P					
Small Organ (Note 1)					P					
Neonatal Cephalic					P					
Adult Cephalic					P					
Cardiac					P					
Trans-esophageal										
Transrectal										
Transvaginal										
Transurethral										
Intravascular										
Peripheral vessel					P					
Laparoscopic										
Musculo-skeletal Conventional					P					
Musculo-skeletal Superficial					P					
Other (specify)										

N = new indication; P = previously cleared by FDA K# 063803, K072786

Additional Comments:

Note 1: For example: breast, testes, thyroid, penis, prostate, etc.

Note 9: For example: vascular, abdominal

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)  
Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use (Per 21 CFR 801.109)

(Division Sign-Off)

Division of Reproductive, Abdominal and  
Radiological Devices

510(k) Number

K090334



Diagnostic Ultrasound Indications for Use Form

510 (k) Number (if known):

Device Name:  
Intended Use:

EC9-4 Curved Array Transducer for use with ACUSON S2000  
Ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation									
	A	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Color Velocity Imaging	Combined (Specify)	Other (Specify)
Ophthalmic										
Fetal		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11
Abdominal		P	P	P		P	P		BMDC	Note 2,3,4,5,6,7,8,10, 11,
Intraoperative Abdominal										
Intraoperative Neurological										
Pediatric										
Small Organ (Note 1)		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11,14
Neonatal Cephalic		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10
Adult Cephalic										
Cardiac										
Trans-esophageal										
Transrectal		P	P	P		P	P		BMDC	Note 2,3,4,5, 6, 7,8,10, 11,14
Transvaginal		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11
Transurethral										
Intravascular										
Peripheral vessel										
Laparoscopic										
Musculo-skeletal Conventional										
Musculo-skeletal Superficial										
Other (specify)										

N = new indication; P = previously cleared by FDA K# 063803, K072786

Additional Comments:

- Note 1 For example: breast, testes, thyroid, penis, prostate, etc.
- Note 2 Ensemble tissue harmonic imaging
- Note 3 SieClear multi-view spatial compounding
- Note 4 Tissue Equalization Technology
- Note 5 3-Scape real-time 3D imaging
- Note 6 Cadence contrast agent imaging
- Note 7 B&W SieScape panoramic imaging
- Note 8 Power SieScape panoramic imaging
- Note 10 Clarify VE vascular enhancement technology
- Note 11 Advanced Sieclear spatial compounding
- Note 14 eSie™ Touch elasticity imaging

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use (Per 21 CFR 801.109)

(Division Sign-Off)

Division of Reproductive, Abdominal and  
Radiological Devices

510(k) Number

Page 16 of 46

Diagnostic Ultrasound Indications for Use Form

510 (k) Number (if known):

Device Name:  
Intended Use:

9L4 Linear Array Transducer for use with ACUSON S2000  
Ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation									
	A	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Color Velocity Imaging	Combined (Specify)	Other (Specify)
Ophthalmic										
Fetal		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11
Abdominal										
Intraoperative Abdominal										
Intraoperative Neurological										
Pediatric		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11
Small Organ (Note 1)		P	P	P		P	P		BMDC	Note 2,3,4,5,6,7,8,10, 11,14
Neonatal Cephalic		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11
Adult Cephalic										
Cardiac		N	N	N		N	N		BMDC	Note 15
Trans-esophageal										
Transrectal										
Transvaginal										
Transurethral										
Intravascular										
Peripheral vessel		P	P	P		P	P		BMDC	Note 2,3,4,5,6, 7,8,10, 11, 14,15
Laparoscopic										
Musculo-skeletal Conventional		P	P	P		P	P		BMDC	Note 2,3,4,5,6,7,8,10, 11, 14
Musculo-skeletal Superficial		P	P	P		P	P		BMDC	Note 2,3,4,5,6,7,8,10, 11, 14
Other (specify)										

N = new indication; P = previously cleared by FDA K# 063085, K072786

Additional Comments:

- Note 1: For example: breast, testes, thyroid, penis, prostate, etc.
- Note 2: Ensemble tissue harmonic imaging
- Note 3: SieClear multi-view spatial compounding
- Note 4: Tissue Equalization Technology
- Note 5: 3-Scape real-time 3D imaging
- Note 6: Cadence contrast agent imaging
- Note 7: B&W SieScape panoramic imaging
- Note 8: Power SieScape panoramic imaging
- Note 10: Clarify VE vascular enhancement technology
- Note 11: Advanced Sieclear spatial compounding
- Note 14: eSie™ Touch elasticity imaging
- Note 15: AHP

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use (Per 21 CFR 801.109)

(Division Sign-Off)

Division of Reproductive, Abdominal and  
Radiological Devices

510(k) Number

K090334

Diagnostic Ultrasound Indications for Use Form

510 (k) Number (if known):

Device Name:

14L5 Multi-D Array Transducer for use with ACUSON S2000

Intended Use:


Ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation									
	A	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Color Velocity Imaging	Combined (Specify)	Other (Specify)
Ophthalmic										
Fetal										
Abdominal										
Intraoperative Abdominal										
Intraoperative Neurological										
Pediatric										
Small Organ (Note 1)		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11, 14
Neonatal Cephalic										
Adult Cephalic										
Cardiac										
Trans-esophageal										
Transrectal										
Transvaginal										
Transurethral										
Intravascular										
Peripheral vessel		P	P	P		P	P		BMDC	Note 2,3,4,5,6, 7,8,10, 11, 14
Laparoscopic										
Musculo-skeletal Conventional		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11, 14
Musculo-skeletal Superficial										
Other (specify)										

N = new indication; P = previously cleared by FDA K# 063085, K072786

Additional Comments:

- Note 1 For example: breast, testes, thyroid, penis, prostate, etc.  
 Note 2 Ensemble tissue harmonic imaging  
 Note 3 SieClear multi-view spatial compounding  
 Note 4 Tissue Equalization Technology  
 Note 5 3-Scape real-time 3D imaging  
 Note 6 Cadence contrast agent imaging  
 Note 7 B&W SieScape panoramic imaging  
 Note 8 Power SieScape panoramic imaging  
 Note 10 Clarify VE vascular enhancement technology  
 Note 11 Advanced Sieclear spatial compounding  
 Note 14 eSie™ Touch elasticity imaging

  
 (Division Sign-Off)  
 Division of Reproductive, Abdominal and  
 Radiological Devices  
 510(k) Number K090334

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)  
 Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use (Per 21 CFR 801.109)

Diagnostic Ultrasound Indications for Use Form

510 (k) Number (if known):

Device Name:

4P1 Phased Array Transducer for use with ACUSON S2000

Intended Use:

Ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation									
	A	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Color Velocity Imaging	Combined (Specify)	Other (Specify)
Ophthalmic										
Fetal		P	P	P	P	P	P		BMDC	Note 2,3,4,5,7,8,10
Abdominal		P	P	P	P	P	P		BMDC	Note 2,3,4,5,7,8,10
Intraoperative Abdominal										
Intraoperative Neurological										
Pediatric										
Small Organ										
Neonatal Cephalic										
Adult Cephalic		P	P	P	P	P	P		BMDC	Note 2,3,4,5,7,8,10
Cardiac		P	P	P	P	P	P		BMDC	Note 2,3,4,5,6,7,8,10
Trans-esophageal										
Transrectal										
Transvaginal										
Transurethral										
Intravascular										
Peripheral vessel										
Laparoscopic										
Musculo-skeletal Conventional										
Musculo-skeletal Superficial										
Other (specify)										

N = new indication; P = previously cleared by FDA K# 063803, K072786

Additional Comments:

Note 2 Ensemble tissue harmonic imaging

Note 3 SieClear multi-view spatial compounding

Note 4 Tissue Equalization Technology

Note 5 3-Scape real-time 3D imaging

Note 6 Cadence contrast agent imaging

Note 7 B&W SieScape panoramic imaging

Note 8 Power SieScape panoramic imaging

Note 10 Clarify VE vascular enhancement technology

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use (Per 21 CFR 801.109)

(Division Sign-Off)

Division of Reproductive, Abdominal and  
Radiological Devices

510(k) Number

K090334

Diagnostic Ultrasound Indications for Use Form

510 (k) Number (if known):


Device Name: 6C2 Curved Array Transducer for use with ACUSON S2000  
Intended Use: Ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation									
	A	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Color Velocity Imaging	Combined (Specify)	Other (Specify)
Ophthalmic										
Fetal		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11
Abdominal		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11, 14
Intraoperative Abdominal										
Intraoperative Neurological										
Pediatric		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11
Small Organ										
Neonatal Cephalic										
Adult Cephalic										
Cardiac										
Trans-esophageal										
Transrectal										
Transvaginal										
Transurethral										
Intravascular										
Peripheral vessel		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11
Laparoscopic										
Musculo-skeletal Conventional										
Musculo-skeletal Superficial										
Other (specify)										

N = new indication; P = previously cleared by FDA K# 063085, K072733

Additional Comments:

- Note 2 Ensemble tissue harmonic imaging
- Note 3 SieClear multi-view spatial compounding
- Note 4 Tissue Equalization Technology
- Note 5 3-Scape real-time 3D imaging
- Note 7 B&W SieScape panoramic imaging
- Note 8 Power SieScape panoramic imaging
- Note 10 Clarify VE vascular enhancement technology
- Note 11 Advanced Sieclear spatial compounding
- Note 14 eSie™ Touch elasticity imaging

  
(Division Sign-Off)  
Division of Reproductive, Abdominal and  
Radiological Devices  
510(k) Number K090334

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)  
Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use (Per 21 CFR 801.109)

Diagnostic Ultrasound Indications for Use Form

510 (k) Number (if known):

Device Name: 4C1 Curved Array Transducer for use with ACUSON S2000  
Intended Use: Ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation									
	A	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Color Velocity Imaging	Combined (Specify)	Other (Specify)
Ophthalmic										
Fetal		P	P	P	P	P	P		BMDC	Note 2,3,4,5,7,8,10, 11
Abdominal		P	P	P	P	P	P		BMDC	Note 2,3,4,5,6,7,8, 10, 11, 14
Intraoperative Abdominal										
Intraoperative Neurological										
Pediatric										
Small Organ		P	P	P	P	P	P		BMDC	
Neonatal Cephalic										
Adult Cephalic										
Cardiac		P	P	P	P	P	P		BMDC	
Trans-esophageal										
Transrectal										
Transvaginal										
Transurethral										
Intravascular										
Peripheral vessel		P	P	P	P	P	P		BMDC	
Laparoscopic										
Musculo-skeletal Conventional										
Musculo-skeletal Superficial										
Other (specify)										

N = new indication; P = previously cleared by FDA K# 063085, K072786, K032114

Additional Comments:

- Note 2: Ensemble tissue harmonic imaging
- Note 3: SieClear multi-view spatial compounding
- Note 4: Tissue Equalization Technology
- Note 5: 3-Scape real-time 3D imaging
- Note 6: Cadence contrast agent imaging
- Note 7: B&W SieScape panoramic imaging
- Note 8: Power SieScape panoramic imaging
- Note 10: Clarify VE vascular enhancement technology
- Note 11: Advanced Sieclear spatial compounding
- Note 14: eSie™ Touch elasticity imaging

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use (Per 21 CFR 801.109)

(Division Sign-Off)

Division of Reproductive, Abdominal and Radiological Devices

510(k) Number

Page 21 of 46

Diagnostic Ultrasound Indications for Use Form

510 (k) Number (if known):

Device Name: 4V1 Phased Array Transducer for use with ACUSON S2000  
Intended Use: Ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation									
	A	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Color Velocity Imaging	Combined (Specify)	Other (Specify)
Ophthalmic										
Fetal		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10
Abdominal		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 14
Intraoperative Abdominal										
Intraoperative Neurological										
Pediatric										
Small Organ										
Neonatal Cephalic										
Adult Cephalic										
Cardiac										
Trans-esophageal										
Transrectal										
Transvaginal										
Transurethral										
Intravascular										
Peripheral vessel										
Laparoscopic										
Musculo-skeletal Conventional										
Musculo-skeletal Superficial										
Other (specify)										

N = new indication; P = previously cleared by FDA K# 063085, K072786

Additional Comments:

- Note 2 Ensemble tissue harmonic imaging
- Note 3 SieClear multi view spatial compounding
- Note 4 Tissue Equalization Technology
- Note 5 3-Scape real-time 3D imaging
- Note 7 B&W SieScape panoramic imaging
- Note 8 Power SieScape panoramic imaging
- Note 10 Clarify VE vascular enhancement technology
- Note 11 Advanced Sieclear spatial compounding
- Note 14 eSie™ Touch elasticity imaging

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)  
Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use (Per 21 CFR 801.109)

(Division Sign-Off)  
Division of Reproductive, Abdominal and Radiological Devices  
510(k) Number K090334

Diagnostic Ultrasound Indications for Use Form

510 (k) Number (if known):

Device Name: 10V4 Phased Array Transducer for use with ACUSON S2000  
Intended Use: Ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation									
	A	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Color Velocity Imaging	Combined (Specify)	Other (Specify)
Ophthalmic										
Fetal		P	P	P	P	P	P		BMDC	Note 2,3,4,5,7,8,10
Abdominal		P	P	P	P	P	P		BMDC	Note 2,3,4,5,7,8,10
Intraoperative Abdominal										
Intraoperative Neurological										
Pediatric		P	P	P	P	P	P		BMDC	Note 2,3,4,5,7,8,10
Small Organ										
Neonatal Cephalic		P	P	P	P	P	P		BMDC	Note 2,3,4,5,7,8,10
Adult Cephalic										
Cardiac		P	P	P	P	P	P		BMDC	Note 3,4
Trans-esophageal										
Transrectal										
Transvaginal										
Transurethral										
Intravascular										
Peripheral vessel		P	P	P	P	P	P		BMDC	Note 2,3,4,5,7,8,10
Laparoscopic										
Musculo-skeletal Conventional										
Musculo-skeletal Superficial										
Other (specify)										


N = new indication; P = previously cleared by FDA K# 063085, K072786

Additional Comments:

- Note 2 Ensemble tissue harmonic imaging
- Note 3 SieClear multi view spatial compounding
- Note 4 Tissue Equalization Technology
- Note 5 3-Scape real-time 3D imaging
- Note 7 B&W SieScape panoramic imaging
- Note 8 Power SieScape panoramic imaging
- Note 10 Clarify VE vascular enhancement technology

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)  
Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use (Per 21 CFR 801.109)

  
(Division Sign-Off)  
Division of Reproductive, Abdominal and  
Radiological Devices  
510(k) Number K090334



Diagnostic Ultrasound Indications for Use Form

510 (k) Number (if known):

Device Name: 14L5 SP Linear Array Transducer for use with ACUSON S2000  
Indications For Use: Diagnostic imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation									
	A	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Color Velocity Imaging	Combined (Specify)	Other (Specify)
Ophthalmic										
Fetal										
Abdominal										
Intraoperative (Note 9)		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10
Intraoperative Neurological		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11
Pediatric										
Small Organ (Note 1)		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11,14
Neonatal Cephalic										
Adult Cephalic										
Cardiac		N	N	N		N	N		BMDC	Note 15
Transesophageal										
Transrectal										
Transvaginal										
Transurethral										
Intravascular										
Peripheral vessel		P	P	P		P	P		BMDC	Note 2,3,4,5,6, 7,8,10, 11,14,15
Laparoscopic										
Musculo-skeletal Conventional		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11,14
Musculo-skeletal Superficial										
Other (specify)										

N = new indication; P = previously cleared by FDA K# 063085, K072786

Additional Comments:

- Note 1 For example: breast, testes, thyroid, penis, prostate, etc.  
 Note 2 Ensemble tissue harmonic imaging  
 Note 3 SieClear multi-view spatial compounding  
 Note 4 Tissue Equalization Technology  
 Note 5 3-Scape real-time 3D imaging  
 Note 6 Cadence contrast agent imaging  
 Note 7 B&W SieScape panoramic imaging  
 Note 8 Power SieScape panoramic imaging  
 Note 9 For example: vascular, abdominal  
 Note 10 Clarify VE vascular enhancement technology  
 Note 11 Advanced Sieclear spatial compounding  
 Note 14 eSie™ Touch elasticity imaging

Note 15 AHP

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation

Prescription Use (Per 21 CFR 804.109)

(Division Sign-Off)

Division of Reproductive, Abdominal and Radiological Devices

510(k) Number

K090334

Diagnostic Ultrasound Indications for Use Form

510 (k) Number (if known):

Device Name: 7CF2 Curved array mechanical 3D transducer for use with ACUSON S2000  
Intended Use: Ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation									
	A	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Color Velocity Imaging	Combined (Specify)	Other (Specify)
Ophthalmic										
Fetal		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11,13
Abdominal		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11, 13
Intraoperative Abdominal										
Intraoperative Neurological										
Pediatric										
Small Organ										
Neonatal Cephalic										
Adult Cephalic										
Cardiac										
Trans-esophageal										
Transrectal										
Transvaginal										
Transurethral										
Intravascular										
Peripheral vessel										
Laparoscopic										
Musculo-skeletal Conventional										
Musculo-skeletal Superficial										
Other (specify)										

N = new indication; P = previously cleared by FDA K# 063803, K072786

Additional Comments:

- Note 2 Ensemble tissue harmonic imaging
- Note 3 SieClear multi-view spatial compounding
- Note 4 Tissue Equalization Technology
- Note 5 3-Scape real-time 3D imaging
- Note 7 B&W SieScape panoramic imaging
- Note 8 Power SieScape panoramic imaging
- Note 10 Clarify VE vascular enhancement technology
- Note 11 Advanced Sieclear spatial compounding
- Note 13 STIC

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

(Division Sign-Off)

Prescription Use (Per 21 CFR 801.409)

Division of Reproductive, Abdominal and Radiological Devices  
510(k) Number K090334

Diagnostic Ultrasound Indications for Use Form

510 (k) Number (if known):

Device Name: 9EVF4 Curved Array Transducer for use with ACUSON S2000  
Intended Use: Ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation									
	A	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Color Velocity Imaging	Combined (Specify)	Other (Specify)
Ophthalmic										
Fetal		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8, 10,11
Abdominal										
Intraoperative Abdominal										
Intraoperative Neurological										
Pediatric										
Small Organ										
Neonatal Cephalic		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8, 10,11
Adult Cephalic										
Cardiac										
Trans-esophageal										
Transrectal										
Transvaginal		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8, 10,11
Transurethral										
Intravascular										
Peripheral vessel										
Laparoscopic										
Musculo-skeletal Conventional										
Musculo-skeletal Superficial										
Other (specify)										

NA = new indication; P = previously cleared by FDA K# 063803, K072736

Additional Comments:

- Note 2 Ensemble tissue harmonic imaging
- Note 3 SieClear multi-view spatial compounding
- Note 4 Tissue Equalization Technology
- Note 5 3-Scape real-time 3D imaging
- Note 7 B&W SieScape panoramic imaging
- Note 8 Power SieScape panoramic imaging
- Note 10 Clarify VE vascular enhancement technology
- Note 11 Advanced Sieclear spatial compounding

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

(Division Sign-Off)

Prescription Use (Per 21 CFR 801.40) Division of Reproductive, Abdominal and

Radiological Devices

510(k) Number

K090334

Diagnostic Ultrasound Indications for Use Form

510 (k) Number (if known):

Device Name:

V5Ms Multiplane TEE Transducer for use with ACUSON S2000

Intended Use:

Ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation									
	A	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Color Velocity Imaging	Combined (Specify)	Other (Specify)
Ophthalmic										
Fetal										
Abdominal										
Intraoperative Abdominal										
Intraoperative Neurological										
Pediatric										
Small Organ										
Neonatal Cephalic										
Adult Cephalic										
Cardiac										
Trans-esophageal		P	P	P	P	P	P		BMDC	
Transrectal										
Transvaginal										
Transurethral										
Intravascular										
Peripheral vessel										
Laparoscopic										
Musculo-skeletal Conventional										
Musculo-skeletal Superficial										
Other (specify)										


N = new indication; P = previously cleared by FDA K# 063803, K072786

Additional Comments: n/a

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use (Per 21 CFR 801.109)

  
(Division Sign-Off)  
Division of Reproductive, Abdominal and  
Radiological Devices  
510(k) Number K090334

Diagnostic Ultrasound Indications for Use Form

510 (k) Number (if known):

Device Name: 17L5HDS Linear Array Transducer for use with ACUSON S2000  
Intended Use: Ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation									
	A	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Color Velocity Imaging	Combined (Specify)	Other (Specify)
Ophthalmic										
Fetal										
Abdominal										
Intraoperative Abdominal										
Intraoperative Neurological										
Pediatric										
Small Organ (Note 1)		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11,14
Neonatal Cephalic										
Adult Cephalic										
Cardiac										
Trans-esophageal										
Transrectal										
Transvaginal										
Transurethral										
Intravascular										
Peripheral vessel		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11,14
Laparoscopic										
Musculo-skeletal Conventional		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11,14
Musculo-skeletal Superficial		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11,14
Other (specify)										

N = new indication; P = previously cleared by FDA K# 063085, K072708


Additional Comments:

- Note 1 For example: breast, testes, thyroid, penis, prostate, etc.
- Note 2 Ensemble tissue harmonic imaging
- Note 3 SieClear multi-view spatial compounding
- Note 4 Tissue Equalization Technology
- Note 5 3-Scape real-time 3D imaging
- Note 7 B&W SieScape panoramic imaging
- Note 8 Power SieScape panoramic imaging
- Note 10 Clarify VE vascular enhancement technology
- Note 11 Advanced Sieclear spatial compounding
- Note 14 eSie™ Touch elasticity imaging

(Division Sign-Off)

Division of Reproductive, Abdominal and Radiological Devices

510(k) Number

  
K090334

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)  
Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use (Per 21 CFR 801.109)

Diagnostic Ultrasound Indications for Use Form

510 (k) Number (if known):

Device Name: 18L6 HD Linear Array Transducer for use with ACUSON S2000  
Intended Use: Ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation									
	A	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Color Velocity Imaging	Combined (Specify)	Other (Specify)
Ophthalmic										
Fetal										
Abdominal										
Intraoperative Abdominal										
Intraoperative Neurological										
Pediatric										
Small Organ (Note 1)		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11,14
Neonatal Cephalic										
Adult Cephalic										
Cardiac		N	N	N		N	N		BMDC	Note 15
Trans-esophageal										
Transrectal										
Transvaginal										
Transurethral										
Intravascular										
Peripheral vessel		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11,14,15
Laparoscopic										
Musculo-skeletal Conventional		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11,14
Musculo-skeletal Superficial		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11,14
Other (specify)										

N = new indication; P = previously cleared by PDA K032142

Additional Comments:

- Note 1 For example: breast, testes, thyroid, penis, prostate, etc.  
 Note 2 Ensemble tissue harmonic imaging  
 Note 3 SieClear multi-view spatial compounding  
 Note 4 Tissue Equalization Technology  
 Note 5 3-Scape real-time 3D imaging  
 Note 7 B&W SieScape panoramic imaging  
 Note 8 Power SieScape panoramic imaging  
 Note 10 Clarify VE vascular enhancement technology  
 Note 11 Advanced Sieclear spatial compounding  
 Note 14 eSie™ Touch elasticity imaging

Note 15 AHP

(Division Sign-Off)

Division of Reproductive, Abdominal and  
Radiological Devices

510(k) Number

K690334

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)  
Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use (Per 21 CFR 801.109)

Diagnostic Ultrasound Indications for Use Form

510 (k) Number (if known):

Device Name:

8V3 Phased Array Transducer for use with ACUSON S2000

Intended Use:

Ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation									
	A	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Color Velocity Imaging	Combined (Specify)	Other (Specify)
Ophthalmic										
Fetal		P	P	P	P	P	P		BMDC	Note 2,3,4,5,7,8,10
Abdominal										
Intraoperative Abdominal										
Intraoperative Neurological										
Pediatric		P	P	P	P	P	P		BMDC	Note 2,3,4,5,7,8,10
Small Organ										
Neonatal Cephalic		P	P	P	P	P	P		BMDC	Note 2,3,4,5,7,8,10
Adult Cephalic										
Cardiac		P	P	P	P	P	P		BMDC	Note 3,4,6
Trans-esophageal										
Transrectal										
Transvaginal										
Transurethral										
Intravascular										
Peripheral vessel										
Laparoscopic										
Musculo-skeletal Conventional										
Musculo-skeletal Superficial										
Other (specify) Neonatal Cardiac		P	P	P	P	P	P		BMDC	Note 3,4,6

N = new indication; P = previously cleared by FDA K# 063085, K072786

Additional Comments:

- Note 2 Ensemble tissue harmonic imaging
- Note 3 SieClear multi-view spatial compounding
- Note 4 Tissue Equalization Technology
- Note 5 3-Scape real-time 3D imaging
- Note 6 Cadence contrast agent imaging
- Note 7 B&W SieScape panoramic imaging
- Note 8 Power SieScape panoramic imaging
- Note 10 Clarify VE vascular enhancement technology

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)  
Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use (Per 21 CFR 801.109)

(Division Sign-Off)

Division of Reproductive, Abdominal and  
Radiological Devices  
510(k) Number K890334

Diagnostic Ultrasound Indications for Use Form

510 (k) Number (if known):

Device Name: 4V1c Phased Array Transducer for use with ACUSON S2000  
Intended Use: Ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation									
	A	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Color Velocity Imaging	Combined (Specify)	Other (Specify)
Ophthalmic										
Fetal		P	P	P	P	P	P		BMDC	Note 2 3 4 5 7 8 10
Abdominal		P	P	P	P	P	P		BMDC	Note 2 3 4 5 7 8 10
Intraoperative Abdominal		P	P	P	P	P	P		BMDC	Note 2 3 4 5 7 8 10
Intraoperative Neurological		P	P	P	P	P	P		BMDC	Note 2 3 4 5 7 8 10
Pediatric		P	P	P	P	P	P		BMDC	Note 2 3 4 5 7 8 10
Small Organ										
Neonatal Cephalic										
Adult Cephalic		P	P	P	P	P	P		BMDC	Note 2 3 4 5 7 8 10
Cardiac		P	P	P	P	P	P		BMDC	Note 2 3 4 5 7 8 10 15
Trans-esophageal										
Transrectal										
Transvaginal										
Transurethral										
Intravascular										
Peripheral vessel		P	P	P	P	P	P		BMDC	Note 2 3 4 5 7 8 10 15
Laparoscopic										
Musculo-skeletal Conventional										
Musculo-skeletal Superficial										
Other (specify) Neonatal Cardiac		P	P	P	P	P	P		BMDC	Note 2 3 4 5 7 8 10

N = new indication; P = previously cleared by FDA K#s 052410, 051139, 041219, 032114, 022567, 063085

Additional Comments:

- Note 2 Ensemble tissue harmonic imaging
- Note 3 SieClear multi-view spatial compounding
- Note 4 Tissue Equalization Technology
- Note 5 3-Scape real-time 3D imaging
- Note 6 Cadence contrast agent imaging
- Note 7 B&W SieScape panoramic imaging
- Note 8 Power SieScape panoramic imaging
- Note 10 Clarify VE vascular enhancement technology
- Note 15 AHP

(Division Sign-Off)

Division of Reproductive, Abdominal and Radiological Devices

510(k) Number

K090334

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use (Per 21 CFR 801.109)



Diagnostic Ultrasound Indications for Use Form

510 (k) Number (if known):

Device Name:

6L3 Transducer for use with ACUSON S2000

Intended Use:

Ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation									
	A	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Color Velocity Imaging	Combined (Specify)	Other (Specify)
Ophthalmic										
Fetal		P	P	P	P	P	P		BMDC	Note 2 3 4 5 7 8 10
Abdominal										
Intraoperative Abdominal		P	P	P	P	P	P		BMDC	Note 2 3 4 5 7 8 10
Intraoperative Neurological		P	P	P	P	P	P		BMDC	Note 2 3 4 5 7 8 10
Pediatric										
Small Organ		P	P	P	P	P	P		BMDC	Note 2 3 4 5 7 8 10
Neonatal Cephalic										
Adult Cephalic										
Cardiac		P	P	P	P	P	P		BMDC	Note 2 3 4 5 7 8 10 15
Trans-esophageal										
Transrectal										
Transvaginal										
Transurethral										
Intravascular										
Peripheral vessel		P	P	P	P	P	P		BMDC	Note 2 3 4 5 7 8 10 15
Laparoscopic										
Musculo-skeletal Conventional		P	P	P	P	P	P		BMDC	Note 2 3 4 5 7 8 10
Musculo-skeletal Superficial		P	P	P	P	P	P		BMDC	Note 2 3 4 5 7 8 10
Other (specify)										

N = new indication; P = previously cleared by FDA K#'s 052410, 051139, 041319, 032114, 022567, 002807, 973767, 063085

Additional Comments:

- Note 2 Ensemble tissue harmonic imaging
- Note 3 SieClear multi-view spatial compounding
- Note 4 Tissue Equalization Technology
- Note 5 3-Scape real-time 3D imaging
- Note 6 Cadence contrast agent imaging
- Note 7 B&W SieScape panoramic imaging
- Note 8 Power SieScape panoramic imaging
- Note 10 Clarify VE vascular enhancement technology
- Note 15 AHP

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use (Per 21 CFR 801.40) (Division Sign-Off)

Division of Reproductive, Abdominal and  
Radiological Devices

510(k) Number

K090834

Diagnostic Ultrasound Indications for Use Form

510 (k) Number (if known):

Device Name:

EV8C4 Transducer for use with ACUSON S2000

Intended Use:

Ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation									
	A	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Color Velocity Imaging	Combined (Specify)	Other (Specify)
Ophthalmic										
Fetal		P	P	P	P	P	P		BMDC	Note 2 3 4 5 7 8 10
Abdominal		P	P	P	P	P	P		BMDC	Note 2 3 4 5 7 8 10
Intraoperative Abdominal										
Intraoperative Neurological										
Pediatric										
Small Organ										
Neonatal Cephalic										
Adult Cephalic										
Cardiac										
Trans-esophageal										
Transrectal										
Transvaginal		P	P	P	P	P	P		BMDC	Note 2 3 4 5 6 7 8 10
Transurethral										
Intravascular										
Peripheral vessel										
Laparoscopic										
Musculo-skeletal Conventional										
Musculo-skeletal Superficial										
Other (specify)										

N = new indication: P = previously cleared by FDA K#'s 052410, 051139, 041319, 032114, 022567, 002807, 973767, 063085

Additional Comments:

- Note 2 Ensemble tissue harmonic imaging
- Note 3 SieClear multi-view spatial compounding
- Note 4 Tissue Equalization Technology
- Note 5 3-Scape real-time 3D imaging
- Note 6 Cadence contrast agent imaging
- Note 7 B&W SieScape panoramic imaging
- Note 8 Power SieScape panoramic imaging
- Note 10 Clarify VE vascular enhancement technology

(Division Sign-off)  
Division of Reproductive, Abdominal and  
Radiological Devices  
510(k) Number K090334

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)  
Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use (Per 21 CFR 801.109)